

**CORRIGENDUM FOR SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF LOWERED ACOUSTIC
DOPPLER CURRENT PROFILER (LADCP).**

TENDER NO. : NCPOR/PS/HSS-514/GT-08

Sl. No.	Tender condition/ Specifications	Revised Tender condition/ Specifications
A	EMD clause	
1	EMD Amount: INR 3,40,000. 00	<p>a) Bidders shall submit amount of Rs. 3,40,000/- (Rupees Three Lakhs Forty Thousand only) towards Earnest Money Deposit (EMD)/Bid Security in any modes acceptable as per GFR.</p> <p>b) EMD EXEMPTION: The bidder seeking EMD exemption must submit the valid supporting document for the applicable EMD exemption (as per GFR's and applicable Government orders). Under the MSE category, only manufacturers for goods for the relevant category are eligible for exemption from EMD.</p>
B	Section IV & Form 3: Technical Specifications and Quality Assurance	
1	Additionally, the LADCP should be capable of operating in both real-time (cabled) and self- contained (internal recording) modes, providing flexibility during various deployment scenarios.	The ADCP should be able to record data in real-time mode (when connected to the CTD) and in self-contained mode (via internal recording). Hence, we want to retain the clause as it is.
2	Clause5: Velocity accuracy: 0.3% of water velocity relative to ADCP ± 0.3 cm/s	Clause 5: Velocity accuracy: 0.5% of water velocity relative to ADCP ± 0.5 cm/s
3	Clause 26: The supplier should provide an instruction manual and supporting documents, including calibration sheets in the original, to validate the quoted specifications.	Clause 26: The supplier should provide an instruction manual and supporting documents, including Calibration/Test sheets in the original, to validate the quoted specifications.